



EASYCAP

EEG Recording Caps and Related Products

EASYCAP GmbH

Am Anger 5

82237 Würthsee

GERMANY

Tel +49 8153 88702-00

www.easycap.de


info@easycap.de

Standard Printed 32Ch actiCAP slim

Cap with electrodes and 64 printed electrode position labels

actiCAP
CCCslim

Electrode Names and Number Labels



Details

Ordering Information

For ordering please give Article Number, Cap Cut, and Size
(Example: *ASPG-32, C-Cut, High Precision, White, 58*):

- Article Number: *ASPG-32*
- Cap Cut: *C-Cut or A-Cut*
- Size (given in cm head circumference):
 - Adult caps: *54, 56, 58, 60, 62, 64* (average male: 58, average female: 56)
 - Children caps: *50* (3-4 years), *52* (5-10 years), *54* (11-14 years)
 - Infant caps: *34, 36* (newborn), *38, 40* (3 months), *42, 44* (7 month), *46, 48* (2 years)

The catalogue-number comprises the cap as described. For further information about accessories or consumables, please visit our website or contact our local distributor.

Cap

Standard: White Subinon Cap with integrated chin belt, with printed name labels.

Sizes 52 – 64 made from High Precision Fabric, Sizes 50 and smaller made from High Comfort Fabric

Options: C-Cut or A-Cut, Size. For further variations, contact us.

Electrodes

Electrodes 1-32 are actiCAP-Slim electrodes, terminating at a splitter box for connection to actiCHamp PLUS. GND (160cm) comes with individual connector. All electrodes are buttoned directly into the cap or can be attached to the skin with washers (= double-sided adhesive rings).

Hints when performing TMS and EEG simultaneously

To minimize the TMS artefacts onto the EEG signal, electrode cables should be led away from the TMS stimulation point. To be able to do so, the electrode cables are not attached to the cap and can be rotated 360° in their cap holes. You may secure a certain cable geometry with the enclosed cable clips and velcro straps.

Depending on the site of the TMS stimulation, it may be advisable to use an electrode as REF in the workspace which is a bit away from the stimulation area.